

Initial Report

Premises/Facility under investigation (name and address)

Alpha Laboratories Incorporated
191 McNaughton Road Unit 101
Vaughan, Ontario L6A 4E2

Type of Premises/Facility

Medical Laboratory and Phlebotomy Services

Date Board of Health became aware of IPAC lapse (yyyy/mm/dd) 2023/03/14	Date of Initial Report posting (yyyy/mm/dd) 2023/03/31
Date of Initial Report update(s) (if applicable) (yyyy/mm/dd)	How the IPAC lapse was identified Complaint

Summary Description of the IPAC Lapse

- Concerns with reprocessing and subsequent re-use of blood collection needle holders/blood tube holders.

IPAC Lapse Investigation	Yes	No	N/A	Please provide further details/steps
Did the IPAC lapse involve a member of a regulatory college?	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
If yes, was the issue referred to the regulatory college?	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
Were any corrective measures recommended and/or implemented?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	

Please provide further details/steps	<p>Corrective measures for Premises/Facility:</p> <ul style="list-style-type: none"> • Cease reprocessing and/or re-use of blood collection needle holders/blood tube holders. • Use disposable single use blood collection needle holders/blood tube holders. • Prepare and use chemical used for disinfection according to Manufacturer's Instructions for Use. • Do not use disinfectants past their expiry date. • Follow indications and established moments for hand hygiene. • Use gloves according to infection prevention and control best practices.
--------------------------------------	---



Infection Prevention and Control Lapse Report

Date any order(s) or directive(s) were issued to the owner/operator (if applicable)

2023-02-17

Initial Report Comments and Contact Information

Any additional Comments: (Please do not include any personal information or personal health information)

York Region Public Health issued an order to the premises on February 17, 2023, outlining corrective measures to be implemented. Follow-up re-inspection was conducted to ensure compliance on February 24, 2023.

If you have any further questions, please contact
Health Connection

Telephone Number
1-800-361-5653

Email Address
Health.inspectors@york.ca

Final Report

Date of Final Report posting (yyyy/mm/dd)

Date any order(s) or directive(s) were issued to the owner/operator (if applicable) (yyyy/mm/dd)

Brief description of corrective measures taken

Final Report Comments and Contact Information

Any Additional Comments: (Please do not include any personal information or personal health information)

If you have any further questions, please contact
Health Connection

Telephone Number
1-800-361-5653

Email Address
Health.inspectors@york.ca