

Initial Report

Premise/Facility under investigation (name and address)

Jing Dental Clinic
49-670 Highway 7 East
Richmond Hill, Ontario
L4B 3P2

Type of Premises/Facility

Dental Clinic

Date Board of Health became aware of IPAC lapse (yyyy/mm/dd) 2023/05/15	Date of Initial Report posting (yyyy/mm/dd) 2023/05/26
Date of Initial Report update(s) (if applicable) (yyyy/mm/dd)	How the IPAC lapse was identified Complaint

Summary Description of the IPAC Lapse

- Concerns with manner of use of air and waterlines after each client and storage of reprocessed dental/medical equipment/devices.

IPAC Lapse Investigation	Yes	No	N/A	Please provide further details/steps
Did the IPAC lapse involve a member of a regulatory college?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
If yes, was the issue referred to the regulatory college?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	Royal College of Dental Surgeons of Ontario
Were any corrective measures recommended and/or implemented?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Please provide further details/steps	Corrective measures for Premises/Facility: <ul style="list-style-type: none"> Activate to discharge air and water for a minimum of 20 seconds after each patient/client use for all devices that contact mucous membranes and are attached to the air or waterlines of the dental unit, including high and low-speed handpieces, and air/water syringe tips. Store sterile dental/medical equipment/devices in their sterile packaging until time of use. 			

Date any order(s) or directive(s) were issued to the owner/operator (if applicable) (yyyy/mm/dd)

Initial Report Comments and Contact Information

Any additional Comments: (Please do not include any personal information or personal health information)

If you have any further questions, please contact
Health Connection

Telephone Number 1-800-361-5653	Email Address Health.inspectors@york.ca
------------------------------------	---

Final Report

Date of Final Report posting (yyyy/mm/dd)
2023/06/01

Date any order(s) or directive(s) were issued to the owner/operator (if applicable) (yyyy/mm/dd)

Correction letter was provided to operator 2023/05/29

Brief description of corrective measures taken



York Region Infection Prevention And Control Lapse Report

The operator demonstrated the corrective measures to activate to discharge air and water for a minimum of 20 seconds after each patient/client use for all devices that contact mucous membranes and are attached to the air or waterlines of the dental unit, including high and low-speed handpieces, and air/water syringe tips and store sterile dental/medical equipment/devices in their sterile packaging until time of use.

Date of all corrective measures were confirmed to have been completed (yyyy/mm/dd)
2023/05/29

Final Report Comments and Contact Information

Any Additional Comments: (Please do not include any personal information or personal health information)

If you have any further questions, please contact

Health Connection

Telephone Number	Email Address
1-800-361-5653	Health.inspectors@york.ca
