

Initial Report
Premises/Facility under investigation (name and address)

Tang Tang Nail
A8 – 4300 Steeles Avenue East
Markham, ON
L3R 0N8

Type of Premises/Facility

Personal Service Setting – Ear Piercing, Body Piercing, Body Piercing Replacement

Date Board of Health became aware of IPAC lapse (yyyy/mm/dd)	Date of Initial Report posting (yyyy/mm/dd)
---	--

2025/08/15

2025/08/21

Date of Initial Report update(s) (if applicable) (yyyy/mm/dd)	How the IPAC lapse was identified
--	--

Complaint

Summary Description of the IPAC Lapse

- Reusable equipment/instrument(s) used in the provision of ear piercing, body piercing and body piercing replacement services were not cleaned, and disinfected or sterilized in accordance with the “Public Health Ontario: Guide to Infection Prevention and Control in Personal Service Settings, 3rd Edition, July 2019”.
- The ear piercing device was not used as per the Manufacturer's Instructions for Use (MIFU).
- Jewelry for body piercing was not maintained as sterile until the point of insertion.
- Needles for body piercing were not pre-sterilized.

IPAC Lapse Investigation	Yes	No	N/A	Please provide further details/steps
Did the IPAC lapse involve a member of a regulatory college?	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
If yes, was the issue referred to the regulatory college?	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	
Were any corrective measures recommended and/or implemented?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Please provide further details/steps	Corrective measures for Premises/Facility: <ul style="list-style-type: none"> • Clean, and disinfect or sterilize reusable equipment/instruments used in the provision of ear piercing, body piercing and body piercing replacement in accordance with the “Public Health Ontario: Guide to Infection Prevention and Control in Personal Service Settings, 3rd Edition, July 2019”. • Use the ear piercing device as per the Manufacturer's Instructions for Use (MIFU). • Use only jewelry for body piercing that is maintained as sterile until the point of insertion. • Use only needles for body piercing that are pre-sterilized. 			

Date any order(s) or directive(s) were issued to the owner/operator (if applicable) (yyyy/mm/dd)

HPPA Section 13 Order issued on 2025/08/14.

Initial Report Comments:



York Region
Infection Prevention and Control Lapse Report

HPPA Section 13 Order was issued on 2025/08/14. The operator was ordered to cease the improper use of ear piercing cartridges, discontinue providing personal services that require the use of sterile equipment, and discontinue use of all equipment used for body piercing replacements.

Any additional Comments: (Please do not include any personal information or personal health information)

If you have any further questions, please contact:

Health Connection

Telephone Number

1-800-361-5653

Email Address

Health.inspectors@york.ca

Final Report

Date of Final Report posting (yyyy/mm/dd)

Date any order(s) or directive(s) were issued to the owner/operator (if applicable) (yyyy/mm/dd)

Brief description of corrective measures taken

Date of all corrective measures were confirmed to have been completed (yyyy/mm/dd)

Final Report Comments and Contact Information

Any Additional Comments: (Please do not include any personal information or personal health information)

If you have any further questions, please contact:

Health Connection

Telephone Number

1-800-361-5653

Email Address

Health.inspectors@york.ca